



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the
periodic maintenance* for the Intoxilyzer 5000 Serial Number
66-COST04 maintained by the Tucson Police Department
Crime Laboratory pursuant to the requirements of the Arizona Department
of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following
dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3,
for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 3/23/06 Cal 1 0.080 AC Cal 2 0.078 AC

Date: 4/20/06 Cal 1 N/A AC Cal 2 N/A AC

Date: N/A Cal 1 AC Cal 2 AC

Between the dates of N/A and N/A
the instrument was operating properly and accurately.


The instrument was taken out of service for repair on 4/20/06.
The instrument was returned to service on N/A.

Jennifer L. Coombs
Jennifer L. Coombs

4/21/06

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 21, day of
April, 2006.



Kathy A. Kinney
Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the
information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 1
file

INTOXILYZER SERVICE RECORD

TS4

Date: April 21, 2006

QAS: Jennifer Coombs 46160

Intoxilyzer Serial Number: 66-005004

Reason For Service: Printer Error during initial diagnostic checks

Date Out of Service: 4/20/06

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:

☒ Sent to Applied Electronics



MEMORANDUM

DATE: April 21, 2006

TO: Whom it may concern

FROM: Jennifer L. Coombs
Intoxilyzer 5000 QAS
Tucson Police Department
Crime Lab
270 S. Stone Avenue
Tucson, AZ 85701-1917
(520) 791-4494

A handwritten signature in cursive script, likely belonging to Jennifer L. Coombs, is positioned to the right of the "FROM:" field.

SUBJECT: Intoxilyzer Out of Service

Intoxilyzer 5000, serial number **66-005004** was taken out of service for repair on **April 20, 2006**.

I cannot testify that it was working properly and accurately between the dates of **March 23, 2006** and **April 20, 2006**.

This instrument was located at **DUI HQ**.